



| | |
|-----------------|---|
| Meter | HI98494 |
| Meter ID | --- |
| Meter SN | M03210033111 |
| Meter Firmware | v1.03 |
| Probe Type | HI7698494 |
| Probe ID | Probe |
| Probe SN | 103240010121 |
| Probe Firmware | v1.01 |
| Cap Model | HI764113-1 |
| Cap SN | 926E788F500104E0 |
| Start Date | 12/14/2022 |
| Lot Name | C07-14m |
| Log Interval | 00:00:15 |
| EC Ref. Temp. | 25 °C |
| EC Temp. Coeff. | 1.9%/ °C |
| TDS Factor | 0.5 |
| Remarks | --- |
| Parameters | pH mVORP DO Saturation DO Concentration Conductivity Temperature |

| GLP pH | Date | Time | Cal. Buff(pH) | Offset | Slope |
|--------|------------|----------|--------------------------------|--------|---------------|
| 1 | 07-02-2023 | 20:07:13 | 4,01 (H), 7,01 (H), 10,01 (H*) | 4,0mV | A: 97% B: 97% |
| 2 | 14-12-2022 | 15:29:57 | 4,01 (H), 7,01 (H), 10,01 (H*) | 6,2mV | A: 97% B: 99% |
| 3 | 09-09-2022 | 10:42:13 | 4,01 (H), 7,01 (H), 10,01 (H*) | 6,7mV | A: 99% B: 99% |
| 4 | 15-06-2022 | 13:46:30 | 4,01 (H), 7,01 (H), 10,01 (H) | 3,0mV | A: 96% B: 97% |
| 5 | ---- | ---- | ---- | ---- | ---- |

| GLP ORP | Date | Time | Point | Offset |
|---------|------------|----------|-------------|--------|
| 1 | 07-02-2023 | 20:04:28 | 470,0 mVORP | 12,2mV |
| 2 | 15-06-2022 | 13:48:50 | 230,0 mVORP | -0,9mV |
| 3 | ---- | ---- | ---- | ---- |
| 4 | ---- | ---- | ---- | ---- |
| 5 | ---- | ---- | ---- | ---- |

| GLP DO | Date | Time | Point |
|--------|------------|----------|---|
| 1 | 14-03-2023 | 15:49:15 | Point1: 100,0 % DO Saturation (H) Point2: 0,0 % DO Saturation (H) |
| 2 | 07-02-2023 | 20:11:36 | Point: 100,0 % DO Saturation (H) |
| 3 | 14-12-2022 | 15:32:32 | Point1: 100,0 % DO Saturation (H) Point2: 0,0 % DO Saturation (H) |
| 4 | ---- | ---- | ---- |
| 5 | ---- | ---- | ---- |

| GLP EC | Date | Time | Point | Offset | Cell Factor |
|--------|------------|----------|-----------------------------|-------------|-------------|
| 1 | 07-02-2023 | 20:09:57 | 1,413 mS/cm Conductivity(H) | 0,000 mS/cm | 4,941/cm |
| 2 | 14-12-2022 | 15:23:52 | 1,413 mS/cm Conductivity(H) | 0,000 mS/cm | 4,934/cm |
| 3 | 09-09-2022 | 11:03:03 | 1,413 mS/cm Conductivity(H) | 0,000 mS/cm | 4,503/cm |
| 4 | 15-06-2022 | 13:48:20 | 5,0 mS/cm Conductivity(H) | 0,000 mS/cm | 4,542/cm |
| 5 | ---- | ---- | ---- | ---- | ---- |

| GLP Temperature | Date | Time | Point |
|-----------------|------|------|-------|
| 1 | ---- | ---- | ---- |
| 2 | ---- | ---- | ---- |
| 3 | ---- | ---- | ---- |
| 4 | ---- | ---- | ---- |
| 5 | ---- | ---- | ---- |

| GLP Pressure | Date | Time | Point |
|--------------|------|------|-------|
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

| Rec | Date | Time | pH | mVORP | %DO | mg/LDO | μS/cm | °C |
|-----|------------|----------|------|-------|-----|--------|-------|-------|
| 1 | 22-03-2023 | 09:32:41 | 7,04 | -79,7 | - | - | 992 | 9,89 |
| 2 | 22-03-2023 | 09:32:56 | 7,04 | -79,7 | - | - | 988 | 9,89 |
| 3 | 22-03-2023 | 09:33:11 | 7,01 | -80,3 | - | - | 976 | 9,95 |
| 4 | 22-03-2023 | 09:33:26 | 7,00 | -81,1 | - | - | 976 | 10,01 |
| 5 | 22-03-2023 | 09:33:41 | 6,99 | -81,9 | - | - | 982 | 10,06 |
| 6 | 22-03-2023 | 09:33:56 | 6,98 | -82,6 | - | - | 988 | 10,11 |
| 7 | 22-03-2023 | 09:34:11 | 6,97 | -83,1 | - | - | 993 | 10,14 |
| 8 | 22-03-2023 | 09:34:26 | 6,97 | -83,5 | - | - | 998 | 10,18 |
| 9 | 22-03-2023 | 09:34:41 | 6,96 | -83,9 | - | - | 1005 | 10,22 |
| 10 | 22-03-2023 | 09:34:56 | 6,95 | -84,3 | - | - | 1022 | 10,24 |
| 11 | 22-03-2023 | 09:35:11 | 6,95 | -84,8 | - | - | 1025 | 10,27 |
| 12 | 22-03-2023 | 09:35:26 | 6,94 | -85,1 | - | - | 1033 | 10,29 |
| 13 | 22-03-2023 | 09:35:41 | 6,94 | -85,5 | - | - | 1036 | 10,30 |
| 14 | 22-03-2023 | 09:35:56 | 6,94 | -85,8 | - | - | 1042 | 10,31 |
| 15 | 22-03-2023 | 09:36:11 | 6,93 | -86,1 | - | - | 1044 | 10,33 |
| 16 | 22-03-2023 | 09:36:26 | 6,93 | -86,4 | - | - | 1048 | 10,34 |
| 17 | 22-03-2023 | 09:36:41 | 6,93 | -86,7 | - | - | 1051 | 10,36 |
| 18 | 22-03-2023 | 09:36:56 | 6,92 | -86,9 | - | - | 1053 | 10,38 |
| 19 | 22-03-2023 | 09:37:11 | 6,92 | -87,1 | - | - | 1053 | 10,38 |
| 20 | 22-03-2023 | 09:37:26 | 6,92 | -87,3 | - | - | 1053 | 10,39 |
| 21 | 22-03-2023 | 09:37:41 | 6,92 | -87,4 | - | - | 1051 | 10,41 |
| 22 | 22-03-2023 | 09:37:56 | 6,92 | -87,5 | - | - | 1051 | 10,41 |
| 23 | 22-03-2023 | 09:38:11 | 6,92 | -87,7 | - | - | 1050 | 10,42 |
| 24 | 22-03-2023 | 09:38:26 | 6,91 | -87,8 | - | - | 1050 | 10,43 |
| 25 | 22-03-2023 | 09:38:41 | 6,91 | -88,0 | - | - | 1050 | 10,43 |
| 26 | 22-03-2023 | 09:38:56 | 6,91 | -88,1 | - | - | 1050 | 10,44 |
| 27 | 22-03-2023 | 09:39:11 | 6,91 | -88,3 | - | - | 1050 | 10,44 |
| 28 | 22-03-2023 | 09:39:26 | 6,91 | -88,5 | - | - | 1050 | 10,45 |
| 29 | 22-03-2023 | 09:39:41 | 6,91 | -88,7 | - | - | 1051 | 10,46 |
| 30 | 22-03-2023 | 09:39:56 | 6,91 | -88,8 | - | - | 1050 | 10,46 |
| 31 | 22-03-2023 | 09:40:11 | 6,91 | -88,9 | - | - | 1049 | 10,47 |
| 32 | 22-03-2023 | 09:40:26 | 6,91 | -89,1 | - | - | 1049 | 10,48 |
| 33 | 22-03-2023 | 09:40:41 | 6,91 | -89,2 | - | - | 1049 | 10,48 |
| 34 | 22-03-2023 | 09:40:56 | 6,91 | -89,3 | - | - | 1049 | 10,48 |
| 35 | 22-03-2023 | 09:41:11 | 6,91 | -89,4 | - | - | 1049 | 10,49 |
| 36 | 22-03-2023 | 09:41:26 | 6,91 | -89,6 | - | - | 1049 | 10,49 |
| 37 | 22-03-2023 | 09:41:41 | 6,91 | -89,8 | - | - | 1050 | 10,50 |

| Rec | Date | Time | pH | mVORP | %DO | mg/LDO | μS/cm | °C |
|-----|------------|----------|------|-------|-----|--------|-------|-------|
| 38 | 22-03-2023 | 09:41:56 | 6,91 | -89,9 | - | - | 1050 | 10,51 |
| 39 | 22-03-2023 | 09:42:11 | 6,91 | -90,0 | - | - | 1050 | 10,52 |
| 40 | 22-03-2023 | 09:42:26 | 6,91 | -90,1 | - | - | 1050 | 10,52 |
| 41 | 22-03-2023 | 09:42:41 | 6,90 | -90,2 | - | - | 1050 | 10,52 |
| 42 | 22-03-2023 | 09:42:56 | 6,90 | -90,3 | - | - | 1050 | 10,51 |
| 43 | 22-03-2023 | 09:43:11 | 6,90 | -90,5 | - | - | 1051 | 10,51 |
| 44 | 22-03-2023 | 09:43:26 | 6,91 | -90,8 | - | - | 1051 | 10,51 |
| 45 | 22-03-2023 | 09:43:41 | 6,91 | -91,0 | - | - | 1050 | 10,51 |
| 46 | 22-03-2023 | 09:43:56 | 6,91 | -91,1 | - | - | 1050 | 10,51 |
| 47 | 22-03-2023 | 09:44:11 | 6,91 | -91,1 | - | - | 1050 | 10,51 |
| 48 | 22-03-2023 | 09:44:26 | 6,91 | -91,3 | - | - | 1050 | 10,52 |
| 49 | 22-03-2023 | 09:44:41 | 6,91 | -91,7 | - | - | 1049 | 10,52 |
| 50 | 22-03-2023 | 09:44:56 | 6,92 | -91,8 | - | - | 1050 | 10,53 |
| 51 | 22-03-2023 | 09:45:11 | 6,92 | -92,0 | - | - | 1049 | 10,53 |
| 52 | 22-03-2023 | 09:45:26 | 6,92 | -92,1 | - | - | 1049 | 10,53 |