



Meter	HI98494
Meter ID	---
Meter SN	M03210033111
Meter Firmware	v1.03
Probe Type	HI7698494
Probe ID	Probe
Probe SN	103240010121
Probe Firmware	v1.01
Cap Model	HI764113-1
Cap SN	926E788F500104E0
Start Date	12/14/2022
Lot Name	A11-15m
Log Interval	00:00:15
EC Ref. Temp.	25 °C
EC Temp. Coeff.	1.9%/ °C
TDS Factor	0.5
Remarks	---
Parameters	pH mVORP DO Saturation DO Concentration Conductivity Temperature

GLP pH	Date	Time	Cal. Buff(pH)	Offset	Slope
1	07-02-2023	20:07:13	4,01 (H), 7,01 (H), 10,01 (H*)	4,0mV	A: 97% B: 97%
2	14-12-2022	15:29:57	4,01 (H), 7,01 (H), 10,01 (H*)	6,2mV	A: 97% B: 99%
3	09-09-2022	10:42:13	4,01 (H), 7,01 (H), 10,01 (H*)	6,7mV	A: 99% B: 99%
4	15-06-2022	13:46:30	4,01 (H), 7,01 (H), 10,01 (H)	3,0mV	A: 96% B: 97%
5	----	----	----	----	----

GLP ORP	Date	Time	Point	Offset
1	07-02-2023	20:04:28	470,0 mVORP	12,2mV
2	15-06-2022	13:48:50	230,0 mVORP	-0,9mV
3	----	----	----	----
4	----	----	----	----
5	----	----	----	----

GLP DO	Date	Time	Point
1	14-03-2023	15:49:15	Point1: 100,0 % DO Saturation (H) Point2: 0,0 % DO Saturation (H)
2	07-02-2023	20:11:36	Point: 100,0 % DO Saturation (H)
3	14-12-2022	15:32:32	Point1: 100,0 % DO Saturation (H) Point2: 0,0 % DO Saturation (H)
4	----	----	----
5	----	----	----

GLP EC	Date	Time	Point	Offset	Cell Factor
1	07-02-2023	20:09:57	1,413 mS/cm Conductivity(H)	0,000 mS/cm	4,941/cm
2	14-12-2022	15:23:52	1,413 mS/cm Conductivity(H)	0,000 mS/cm	4,934/cm
3	09-09-2022	11:03:03	1,413 mS/cm Conductivity(H)	0,000 mS/cm	4,503/cm
4	15-06-2022	13:48:20	5,0 mS/cm Conductivity(H)	0,000 mS/cm	4,542/cm
5	----	----	----	----	----

GLP Temperature	Date	Time	Point
1	----	----	----
2	----	----	----
3	----	----	----
4	----	----	----
5	----	----	----

GLP Pressure	Date	Time	Point

Rec	Date	Time	pH	mVORP	%DO	mg/LDO	μS/cm	°C
1	22-03-2023	08:36:23	7,29	-101,7	-	-	805	11,07
2	22-03-2023	08:36:38	7,29	-103,7	-	-	811	11,09
3	22-03-2023	08:36:53	7,30	-106,1	-	-	816	11,11
4	22-03-2023	08:37:08	7,30	-108,4	-	-	819	11,13
5	22-03-2023	08:37:23	7,31	-110,6	-	-	820	11,14
6	22-03-2023	08:37:38	7,31	-112,7	-	-	823	11,15
7	22-03-2023	08:37:53	7,31	-114,5	-	-	825	11,16
8	22-03-2023	08:38:08	7,31	-116,1	-	-	824	11,16
9	22-03-2023	08:38:23	7,31	-117,8	-	-	818	11,17
10	22-03-2023	08:38:38	7,31	-119,4	-	-	808	11,16
11	22-03-2023	08:38:53	7,30	-121,0	-	-	795	11,16
12	22-03-2023	08:39:08	7,30	-122,3	-	-	784	11,15
13	22-03-2023	08:39:23	7,30	-123,1	-	-	774	11,14
14	22-03-2023	08:39:38	7,29	-123,6	-	-	773	11,13
15	22-03-2023	08:39:53	7,29	-123,8	-	-	771	11,13
16	22-03-2023	08:40:08	7,28	-124,1	-	-	767	11,13
17	22-03-2023	08:40:23	7,28	-124,2	-	-	764	11,14
18	22-03-2023	08:40:38	7,28	-124,2	-	-	761	11,15
19	22-03-2023	08:40:53	7,27	-124,3	-	-	759	11,17
20	22-03-2023	08:41:08	7,27	-124,3	-	-	756	11,18
21	22-03-2023	08:41:23	7,27	-124,2	-	-	754	11,18
22	22-03-2023	08:41:38	7,26	-124,1	-	-	752	11,18
23	22-03-2023	08:41:53	7,26	-124,1	-	-	750	11,19
24	22-03-2023	08:42:08	7,26	-123,9	-	-	749	11,19
25	22-03-2023	08:42:23	7,25	-123,8	-	-	746	11,19
26	22-03-2023	08:42:38	7,25	-123,6	-	-	745	11,20
27	22-03-2023	08:42:53	7,25	-123,4	-	-	742	11,20
28	22-03-2023	08:43:08	7,24	-123,3	-	-	740	11,21
29	22-03-2023	08:43:23	7,24	-123,0	-	-	739	11,22
30	22-03-2023	08:43:38	7,24	-122,8	-	-	738	11,22
31	22-03-2023	08:43:53	7,23	-122,5	-	-	738	11,23
32	22-03-2023	08:44:08	7,23	-122,3	-	-	736	11,24
33	22-03-2023	08:44:23	7,23	-122,0	-	-	735	11,25
34	22-03-2023	08:44:38	7,23	-121,8	-	-	734	11,26
35	22-03-2023	08:44:53	7,22	-121,6	-	-	734	11,27
36	22-03-2023	08:45:08	7,22	-121,3	-	-	732	11,27
37	22-03-2023	08:45:23	7,22	-121,2	-	-	730	11,27

Rec	Date	Time	pH	mVORP	%DO	mg/LDO	μS/cm	°C
38	22-03-2023	08:45:38	7,22	-121,0	-	-	730	11,27
39	22-03-2023	08:45:53	7,21	-120,8	-	-	729	11,27
40	22-03-2023	08:46:08	7,21	-120,5	-	-	728	11,28
41	22-03-2023	08:46:23	7,21	-120,3	-	-	725	11,29
42	22-03-2023	08:46:38	7,21	-120,1	-	-	730	11,29
43	22-03-2023	08:46:53	7,20	-119,9	-	-	729	11,30
44	22-03-2023	08:47:08	7,20	-119,7	-	-	727	11,31
45	22-03-2023	08:47:23	7,20	-119,5	-	-	726	11,31
46	22-03-2023	08:47:38	7,20	-119,2	-	-	722	11,31
47	22-03-2023	08:47:53	7,19	-118,9	-	-	722	11,31
48	22-03-2023	08:48:08	7,19	-118,4	-	-	720	11,31
49	22-03-2023	08:48:23	7,18	-118,2	-	-	719	11,32
50	22-03-2023	08:48:38	7,18	-118,0	-	-	718	11,32
51	22-03-2023	08:48:53	7,18	-117,9	-	-	717	11,33
52	22-03-2023	08:49:08	7,18	-117,8	-	-	717	11,33