



Meter	HI98494
Meter ID	---
Meter SN	M03210033111
Meter Firmware	v1.03
Probe Type	HI7698494
Probe ID	Probe
Probe SN	103240010121
Probe Firmware	v1.01
Cap Model	HI764113-1
Cap SN	926E788F500104E0
Start Date	12/14/2022
Lot Name	B09-15m
Log Interval	00:00:15
EC Ref. Temp.	25 °C
EC Temp. Coeff.	1.9%/ °C
TDS Factor	0.5
Remarks	---
Parameters	pH mVORP DO Saturation DO Concentration Conductivity Temperature

GLP pH	Date	Time	Cal. Buff(pH)	Offset	Slope
1	07-02-2023	20:07:13	4,01 (H), 7,01 (H), 10,01 (H*)	4,0mV	A: 97% B: 97%
2	14-12-2022	15:29:57	4,01 (H), 7,01 (H), 10,01 (H*)	6,2mV	A: 97% B: 99%
3	09-09-2022	10:42:13	4,01 (H), 7,01 (H), 10,01 (H*)	6,7mV	A: 99% B: 99%
4	15-06-2022	13:46:30	4,01 (H), 7,01 (H), 10,01 (H)	3,0mV	A: 96% B: 97%
5	----	----	----	----	----

GLP ORP	Date	Time	Point	Offset
1	07-02-2023	20:04:28	470,0 mVORP	12,2mV
2	15-06-2022	13:48:50	230,0 mVORP	-0,9mV
3	----	----	----	----
4	----	----	----	----
5	----	----	----	----

GLP DO	Date	Time	Point
1	14-03-2023	15:49:15	Point1: 100,0 % DO Saturation (H) Point2: 0,0 % DO Saturation (H)
2	07-02-2023	20:11:36	Point: 100,0 % DO Saturation (H)
3	14-12-2022	15:32:32	Point1: 100,0 % DO Saturation (H) Point2: 0,0 % DO Saturation (H)
4	----	----	----
5	----	----	----

GLP EC	Date	Time	Point	Offset	Cell Factor
1	07-02-2023	20:09:57	1,413 mS/cm Conductivity(H)	0,000 mS/cm	4,941/cm
2	14-12-2022	15:23:52	1,413 mS/cm Conductivity(H)	0,000 mS/cm	4,934/cm
3	09-09-2022	11:03:03	1,413 mS/cm Conductivity(H)	0,000 mS/cm	4,503/cm
4	15-06-2022	13:48:20	5,0 mS/cm Conductivity(H)	0,000 mS/cm	4,542/cm
5	----	----	----	----	----

GLP Temperature	Date	Time	Point
1	----	----	----
2	----	----	----
3	----	----	----
4	----	----	----
5	----	----	----

GLP Pressure	Date	Time	Point

Rec	Date	Time	pH	mVORP	%DO	mg/LDO	μS/cm	°C
1	22-03-2023	08:11:37	7,45	-105,4	7,9	0,81	1229	10,46
2	22-03-2023	08:11:52	7,46	-113,2	3,6	0,37	1204	10,49
3	22-03-2023	08:12:07	7,47	-118,7	1,4	0,14	1076	10,51
4	22-03-2023	08:12:22	7,48	-122,9	-	-	1047	10,53
5	22-03-2023	08:12:37	7,49	-126,0	-	-	1034	10,54
6	22-03-2023	08:12:52	7,49	-128,9	-	-	1005	10,55
7	22-03-2023	08:13:07	7,49	-131,2	-	-	981	10,56
8	22-03-2023	08:13:22	7,50	-133,1	-	-	978	10,57
9	22-03-2023	08:13:37	7,50	-134,9	-	-	968	10,58
10	22-03-2023	08:13:52	7,50	-136,3	-	-	966	10,58
11	22-03-2023	08:14:07	7,51	-137,6	-	-	964	10,58
12	22-03-2023	08:14:22	7,51	-138,8	-	-	960	10,58
13	22-03-2023	08:14:37	7,51	-139,8	-	-	961	10,59
14	22-03-2023	08:14:52	7,51	-140,6	-	-	959	10,59
15	22-03-2023	08:15:07	7,51	-141,5	-	-	958	10,60
16	22-03-2023	08:15:22	7,51	-142,3	-	-	958	10,60
17	22-03-2023	08:15:37	7,51	-143,0	-	-	958	10,60
18	22-03-2023	08:15:52	7,52	-143,7	-	-	956	10,60
19	22-03-2023	08:16:07	7,52	-144,3	-	-	954	10,60
20	22-03-2023	08:16:22	7,52	-144,9	-	-	951	10,60
21	22-03-2023	08:16:37	7,52	-145,4	-	-	950	10,60
22	22-03-2023	08:16:52	7,52	-145,9	-	-	949	10,60
23	22-03-2023	08:17:07	7,52	-146,3	-	-	948	10,61
24	22-03-2023	08:17:22	7,52	-146,8	-	-	945	10,61
25	22-03-2023	08:17:37	7,52	-147,1	-	-	944	10,61
26	22-03-2023	08:17:52	7,52	-147,4	-	-	944	10,61
27	22-03-2023	08:18:07	7,52	-147,8	-	-	942	10,61
28	22-03-2023	08:18:22	7,52	-148,2	-	-	942	10,60
29	22-03-2023	08:18:37	7,53	-148,7	-	-	942	10,59
30	22-03-2023	08:18:52	7,53	-149,4	-	-	939	10,59
31	22-03-2023	08:19:07	7,53	-150,6	-	-	949	10,57
32	22-03-2023	08:19:22	7,53	-151,2	-	-	537	10,57
33	22-03-2023	08:19:37	7,53	-151,9	-	-	776	10,57
34	22-03-2023	08:19:52	7,53	-152,6	-	-	747	10,57
35	22-03-2023	08:20:07	7,53	-153,2	-	-	603	10,57
36	22-03-2023	08:20:22	7,54	-153,7	-	-	950	10,43
37	22-03-2023	08:20:37	7,54	-154,1	-	-	950	10,36

Rec	Date	Time	pH	mVORP	%DO	mg/LDO	μS/cm	°C
38	22-03-2023	08:20:52	7,54	-154,3	-	-	948	10,38
39	22-03-2023	08:21:07	7,53	-153,9	-	-	950	10,41
40	22-03-2023	08:21:22	7,53	-153,5	-	-	951	10,42
41	22-03-2023	08:21:37	7,53	-153,3	-	-	953	10,44
42	22-03-2023	08:21:52	7,53	-153,2	-	-	952	10,46
43	22-03-2023	08:22:07	7,53	-153,2	-	-	951	10,47
44	22-03-2023	08:22:22	7,53	-153,1	-	-	951	10,49
45	22-03-2023	08:22:37	7,53	-153,1	-	-	951	10,51
46	22-03-2023	08:22:52	7,53	-153,1	-	-	951	10,53
47	22-03-2023	08:23:07	7,53	-153,1	-	-	949	10,54
48	22-03-2023	08:23:22	7,53	-153,1	-	-	949	10,55
49	22-03-2023	08:23:37	7,53	-153,2	-	-	948	10,56
50	22-03-2023	08:23:52	7,53	-153,2	-	-	946	10,57